

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092625\
 Data File : VX047852.D
 Acq On : 26 Sep 2025 13:01
 Operator : JC/MD
 Sample : Q3127-02DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20250915DL

Quant Time: Sep 26 23:12:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	121604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	248889	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	250346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	121706	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	94625	48.886	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.780%
35) Dibromofluoromethane	5.379	113	80258	48.082	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.160%
50) Toluene-d8	8.635	98	279266	48.352	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.700%
62) 4-Bromofluorobenzene	11.067	95	113991	52.003	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.000%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.110	130	74439	41.563	ug/l	95
64) Tetrachloroethene	9.256	164	4832	2.961	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092625\
Data File : VX047852.D
Acq On : 26 Sep 2025 13:01
Operator : JC/MD
Sample : Q3127-02DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20250915DL

Quant Time: Sep 26 23:12:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

