

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092625\
 Data File : VX047854.D
 Acq On : 26 Sep 2025 13:43
 Operator : JC/MD
 Sample : Q3156-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT190D1-20250918DL

Quant Time: Sep 26 23:14:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	129371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	270470	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	274305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	134213	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	109732	53.287	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	106.580%	
35) Dibromofluoromethane	5.373	113	90161	49.705	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.420%	
50) Toluene-d8	8.635	98	303928	48.423	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	96.840%	
62) 4-Bromofluorobenzene	11.061	95	128407	53.906	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	107.820%	
Target Compounds						
44) Trichloroethene	7.110	130	123155	63.276	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092625\
Data File : VX047854.D
Acq On : 26 Sep 2025 13:43
Operator : JC/MD
Sample : Q3156-04DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT190D1-20250918DL

Quant Time: Sep 26 23:14:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

