

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092625\
 Data File : VX047862.D
 Acq On : 26 Sep 2025 16:31
 Operator : JC/MD
 Sample : Q3156-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB03-20250918

Quant Time: Sep 26 23:17:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	117431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	245049	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	248321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	123059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	95765	51.233	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.460%
35) Dibromofluoromethane	5.373	113	80276	48.847	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.700%
50) Toluene-d8	8.634	98	275851	48.509	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.020%
62) 4-Bromofluorobenzene	11.067	95	114349	52.984	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.960%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	13534	16.652	ug/l	100
29) Tetrahydrofuran	5.013	42	3947	6.472	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092625\
Data File : VX047862.D
Acq On : 26 Sep 2025 16:31
Operator : JC/MD
Sample : Q3156-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB03-20250918

Quant Time: Sep 26 23:17:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

