

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092718\
 Data File : VX004771.D
 Acq On : 27 Sep 2018 13:56
 Operator : JC/MD
 Sample : J5049-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-40-SEP18

Quant Time: Sep 28 12:49:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	265713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	424606	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	406723	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	228392	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	193667	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
35) Dibromofluoromethane	5.50	113	165880	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
50) Toluene-d8	8.72	98	612809	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
62) 4-Bromofluorobenzene	11.14	95	218363	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
Target Compounds						
16) Acetone	2.45	43	251004	113.268	ug/l	98
25) 2-Butanone	4.70	43	65630	20.092	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092718\
Data File : VX004771.D
Acq On : 27 Sep 2018 13:56
Operator : JC/MD
Sample : J5049-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-40-SEP18

Quant Time: Sep 28 12:49:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

