

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092718\
 Data File : VX004775.D
 Acq On : 27 Sep 2018 15:43
 Operator : JC/MD
 Sample : J5048-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-123-SEP18

Quant Time: Sep 28 12:57:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251194	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	409085	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	397981	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	222588	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	190542	53.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.20%	
35) Dibromofluoromethane	5.50	113	160856	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
50) Toluene-d8	8.72	98	591512	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
62) 4-Bromofluorobenzene	11.14	95	213405	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
Target Compounds						
4) Vinyl Chloride	1.40	62	3021	0.775	ug/l #	31
9) 1,1,2-Trichlorotrifluoroet	2.39	101	1543	0.561	ug/l	93
12) 1,1-Dichloroethene	2.37	96	6292	2.312	ug/l	95
24) 1,1-Dichloroethane	3.70	63	49808	8.350	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	203235	61.060	ug/l	95
44) Trichloroethene	7.21	130	113872	31.279	ug/l	93
64) Tetrachloroethene	9.34	164	138587	35.381	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092718\
Data File : VX004775.D
Acq On : 27 Sep 2018 15:43
Operator : JC/MD
Sample : J5048-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-123-SEP18

Quant Time: Sep 28 12:57:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

