

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092718\
 Data File : VX004780.D
 Acq On : 27 Sep 2018 17:57
 Operator : JC/MD
 Sample : J5048-12 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-144-SEP18

Quant Time: Sep 28 07:16:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	236310	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	376475	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	364008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199549	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	177807	52.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.34%	
35) Dibromofluoromethane	5.50	113	148484	46.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.08%	
50) Toluene-d8	8.71	98	541999	51.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.20%	
62) 4-Bromofluorobenzene	11.14	95	193281	50.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.16%	
Target Compounds						
4) Vinyl Chloride	1.40	62	55808	15.222	ug/l	96
12) 1,1-Dichloroethene	2.37	96	2199	0.859	ug/l	99
16) Acetone	2.45	43	8459	4.292	ug/l	90
27) cis-1,2-Dichloroethene	4.60	96	212796	67.959	ug/l	95
44) Trichloroethene	7.21	130	82768	24.704	ug/l	89
64) Tetrachloroethene	9.34	164	2295	0.641	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092718\
Data File : VX004780.D
Acq On : 27 Sep 2018 17:57
Operator : JC/MD
Sample : J5048-12 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-144-SEP18

Quant Time: Sep 28 07:16:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

