

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092718\
 Data File : VX004782.D
 Acq On : 27 Sep 2018 18:50
 Operator : JC/MD
 Sample : J5048-03 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-57-SEP18

Quant Time: Sep 28 13:17:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226503	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	372619	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	374020	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	223061	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	169875	52.50	ug/l	0.00
Spiked Amount			Recovery	=	105.00%	
35) Dibromofluoromethane	5.50	113	142628	45.16	ug/l	0.00
Spiked Amount			Recovery	=	90.32%	
50) Toluene-d8	8.72	98	549983	52.39	ug/l	0.00
Spiked Amount			Recovery	=	104.78%	
62) 4-Bromofluorobenzene	11.14	95	215140	56.88	ug/l	0.00
Spiked Amount			Recovery	=	113.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	3998	1.138	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	2.38	101	11343	4.577	ug/l	97
12) 1,1-Dichloroethene	2.37	96	2566	1.046	ug/l #	68
16) Acetone	2.45	43	427738	226.436	ug/l	95
24) 1,1-Dichloroethane	3.69	63	18233	3.390	ug/l	97
25) 2-Butanone	4.71	43	6451	2.317	ug/l	90
27) cis-1,2-Dichloroethene	4.60	96	66987	22.319	ug/l	97
30) Chloroform	5.23	83	3266	0.642	ug/l	86
32) 1,1,1-Trichloroethane	5.49	97	27532	6.632	ug/l	99
37) Ethyl Acetate	4.86	43	189829	33.191	ug/l	98
44) Trichloroethene	7.21	130	72060	21.731	ug/l	93
51) 4-Methyl-2-Pentanone	8.65	43	63859	12.286	ug/l	94
64) Tetrachloroethene	9.34	164	57879	15.723	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092718\
Data File : VX004782.D
Acq On : 27 Sep 2018 18:50
Operator : JC/MD
Sample : J5048-03 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-57-SEP18

Quant Time: Sep 28 13:17:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

