

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092718\
 Data File : VX004783.D
 Acq On : 27 Sep 2018 19:17
 Operator : JC/MD
 Sample : J5048-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CMT-22-6-SEP18

Quant Time: Sep 28 13:19:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	236939	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	387996	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	383841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	220713	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	182774	54.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.00%	
35) Dibromofluoromethane	5.51	113	153405	46.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.30%	
50) Toluene-d8	8.72	98	563652	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
62) 4-Bromofluorobenzene	11.14	95	208735	53.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.00%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	10715	3.413	ug/l	94
44) Trichloroethene	7.21	130	10926	3.164	ug/l	99
64) Tetrachloroethene	9.34	164	1909	0.505	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092718\
Data File : VX004783.D
Acq On : 27 Sep 2018 19:17
Operator : JC/MD
Sample : J5048-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CMT-22-6-SEP18

Quant Time: Sep 28 13:19:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

