

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092718\
 Data File : VX004804.D
 Acq On : 28 Sep 2018 05:31
 Operator : JC/MD
 Sample : J5033-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPSI-TT-MW301I-20180918

Quant Time: Sep 28 09:56:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	225258	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	368685	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	360918	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199292	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	180525	56.10	ug/l	0.00
Spiked Amount			Recovery	=	112.20%	
35) Dibromofluoromethane	5.50	113	149057	47.70	ug/l	0.00
Spiked Amount			Recovery	=	95.40%	
50) Toluene-d8	8.72	98	539180	51.90	ug/l	0.00
Spiked Amount			Recovery	=	103.80%	
62) 4-Bromofluorobenzene	11.14	95	190952	51.03	ug/l	0.00
Spiked Amount			Recovery	=	102.06%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092718\
Data File : VX004804.D
Acq On : 28 Sep 2018 05:31
Operator : JC/MD
Sample : J5033-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPSI-TT-MW301I-20180918

Quant Time: Sep 28 09:56:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

