

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092718\
 Data File : VX004806.D
 Acq On : 28 Sep 2018 06:24
 Operator : JC/MD
 Sample : J5033-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-SW4004-20180918

Quant Time: Sep 28 09:57:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	231919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	363671	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	361851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	203596	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	182166	54.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.98%	
35) Dibromofluoromethane	5.50	113	149514	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
50) Toluene-d8	8.71	98	539536	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
62) 4-Bromofluorobenzene	11.14	95	193533	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
Target Compounds						
16) Acetone	2.45	43	40712	21.049	ug/l	94
25) 2-Butanone	4.70	43	5193	1.821	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092718\
Data File : VX004806.D
Acq On : 28 Sep 2018 06:24
Operator : JC/MD
Sample : J5033-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-SW4004-20180918

Quant Time: Sep 28 09:57:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

