

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092718\
 Data File : VX004807.D
 Acq On : 28 Sep 2018 06:51
 Operator : JC/MD
 Sample : J5101-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C-1329

Quant Time: Sep 28 08:39:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	237880	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	382455	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	376054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	214990	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	189310	55.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.42%	
35) Dibromofluoromethane	5.50	113	157575	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
50) Toluene-d8	8.72	98	558404	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
62) 4-Bromofluorobenzene	11.14	95	203165	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
Target Compounds						
11) Tert butyl alcohol	3.07	59	27385	28.822	ug/l	99
16) Acetone	2.45	43	10909	5.499	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092718\
Data File : VX004807.D
Acq On : 28 Sep 2018 06:51
Operator : JC/MD
Sample : J5101-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C-1329

Quant Time: Sep 28 08:39:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

