

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092718\
 Data File : VX004810.D
 Acq On : 28 Sep 2018 08:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 28 08:45:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.682	0.843	-23.6#	85	0.00
3 P	Chloromethane	0.827	0.848	-2.5	78	0.00
4 C	Vinyl Chloride	0.776	0.885	-14.0#	86	0.00
5 T	Bromomethane	0.460	0.434	5.7	74	0.00
6 T	Chloroethane	0.495	0.494	0.2	88	0.00
7 T	Trichlorofluoromethane	0.960	1.089	-13.4	83	0.00
8 T	Diethyl Ether	0.417	0.436	-4.6	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.547	0.612	-11.9	94	0.00
10 T	Methyl Iodide	0.732	0.423	42.2#	40#	0.00
11 T	Tert butyl alcohol	0.200	0.209	-4.5	84	0.00
12 CM	1,1-Dichloroethene	0.542	0.590	-8.9#	91	0.00
13 T	Acrolein	0.142	0.117	17.6	69	0.00
14 T	Allyl chloride	1.182	1.256	-6.3	83	0.00
15 T	Acrylonitrile	0.440	0.474	-7.7	84	0.00
16 T	Acetone	0.417	0.409	1.9	76	0.00
17 T	Carbon Disulfide	1.628	1.760	-8.1	85	0.00
18 T	Methyl Acetate	1.056	1.199	-13.5	92	0.00
19 T	Methyl tert-butyl Ether	1.869	2.114	-13.1	93	0.00
20 T	Methylene Chloride	0.670	0.679	-1.3	80	0.00
21 T	trans-1,2-Dichloroethene	0.593	0.646	-8.9	85	0.00
22 T	Diisopropyl ether	1.991	2.214	-11.2	87	0.00
23 T	Vinyl Acetate	1.822	2.002	-9.9	83	0.00
24 P	1,1-Dichloroethane	1.187	1.247	-5.1	82	0.00
25 T	2-Butanone	0.615	0.636	-3.4	82	0.00
26 T	2,2-Dichloropropane	0.827	0.559	32.4#	55	0.00
27 T	cis-1,2-Dichloroethene	0.663	0.683	-3.0	80	0.00
28 T	Bromochloromethane	0.559	0.535	4.3	80	0.00
29 T	Tetrahydrofuran	0.382	0.411	-7.6	82	0.00
30 C	Chloroform	1.122	1.132	-0.9#	85	0.00
31 T	Cyclohexane	0.909	1.072	-17.9	86	0.00
32 T	1,1,1-Trichloroethane	0.916	0.977	-6.7	86	0.00
33 S	1,2-Dichloroethane-d4	0.714	0.694	2.8	74	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.424	0.381	10.1	71	0.00
36 T	1,1-Dichloropropene	0.578	0.578	0.0	85	0.00
37 T	Ethyl Acetate	0.767	0.761	0.8	81	0.00
38 T	Carbon Tetrachloride	0.608	0.584	3.9	81	0.00
39 T	Methylcyclohexane	0.590	0.680	-15.3	97	0.00
40 TM	Benzene	1.810	1.753	3.1	83	0.00
41 T	Methacrylonitrile	0.426	0.430	-0.9	80	0.00
42 TM	1,2-Dichloroethane	0.640	0.629	1.7	83	0.00
43 T	Isopropyl Acetate	1.108	1.148	-3.6	80	0.00
44 TM	Trichloroethene	0.445	0.464	-4.3	95	0.00
45 C	1,2-Dichloropropane	0.444	0.475	-7.0#	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092718\
 Data File : VX004810.D
 Acq On : 28 Sep 2018 08:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 28 08:45:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.279	0.304	-9.0	97	0.00
47 T	Bromodichloromethane	0.522	0.575	-10.2	96	0.00
48 T	Methyl methacrylate	0.503	0.607	-20.7#	98	0.00
49 T	1,4-Dioxane	0.012	0.013	-8.3	83	0.00
50 S	Toluene-d8	1.409	1.459	-3.5	91	0.00
51 T	4-Methyl-2-Pentanone	0.697	0.809	-16.1	95	0.00
52 CM	Toluene	0.995	1.091	-9.6#	90	0.00
53 T	t-1,3-Dichloropropene	0.590	0.607	-2.9	80	0.00
54 T	cis-1,3-Dichloropropene	0.618	0.669	-8.3	90	0.00
55 T	1,1,2-Trichloroethane	0.436	0.452	-3.7	86	0.00
56 T	Ethyl methacrylate	0.628	0.724	-15.3	90	0.00
57 T	1,3-Dichloropropane	0.725	0.761	-5.0	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.331	0.395	-19.3	92	0.00
59 T	2-Hexanone	0.581	0.644	-10.8	85	0.00
60 T	Dibromochloromethane	0.436	0.465	-6.7	81	0.00
61 T	1,2-Dibromoethane	0.467	0.471	-0.9	78	0.00
62 S	4-Bromofluorobenzene	0.508	0.537	-5.7	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.492	0.475	3.5	83	0.00
65 PM	Chlorobenzene	1.282	1.265	1.3	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.451	0.461	-2.2	92	0.00
67 C	Ethyl Benzene	2.036	2.189	-7.5#	96	0.00
68 T	m/p-Xylenes	0.795	0.847	-6.5	91	0.00
69 T	o-Xylene	0.731	0.820	-12.2	97	0.00
70 T	Styrene	1.227	1.385	-12.9	96	0.00
71 P	Bromoform	0.386	0.398	-3.1	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.152	3.602	-14.3	97	0.00
74 T	N-amyl acetate	1.531	1.814	-18.5	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.283	1.291	-0.6	97	0.00
76 T	1,2,3-Trichloropropane	1.121	1.154	-2.9	95	0.00
77 T	Bromobenzene	0.915	0.967	-5.7	96	0.00
78 T	n-propylbenzene	3.616	4.181	-15.6	98	0.00
79 T	2-Chlorotoluene	2.254	2.478	-9.9	96	0.00
80 T	1,3,5-Trimethylbenzene	2.664	3.063	-15.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.367	0.322	12.3	79	0.00
82 T	4-Chlorotoluene	2.653	2.931	-10.5	96	0.00
83 T	tert-Butylbenzene	2.595	3.008	-15.9	99	0.00
84 T	1,2,4-Trimethylbenzene	2.728	3.186	-16.8	97	0.00
85 T	sec-Butylbenzene	3.161	3.641	-15.2	99	0.00
86 T	p-Isopropyltoluene	2.863	3.268	-14.1	97	0.00
87 T	1,3-Dichlorobenzene	1.709	1.771	-3.6	95	0.00
88 T	1,4-Dichlorobenzene	1.765	1.798	-1.9	95	0.00
89 T	n-Butylbenzene	2.601	2.902	-11.6	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092718\
Data File : VX004810.D
Acq On : 28 Sep 2018 08:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 28 08:45:31 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.477	0.526	-10.3	105	0.00
91 T	1,2-Dichlorobenzene	1.786	1.805	-1.1	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.282	0.307	-8.9	102	0.00
93 T	1,2,4-Trichlorobenzene	1.253	1.347	-7.5	96	0.00
94 T	Hexachlorobutadiene	0.650	0.635	2.3	95	0.00
95 T	Naphthalene	3.727	4.402	-18.1	98	0.00
96 T	1,2,3-Trichlorobenzene	1.294	1.380	-6.6	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6