

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092718\
 Data File : VX004804.D
 Acq On : 28 Sep 2018 05:31
 Operator : JC/MD
 Sample : J5033-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPSI-TT-MW301I-20180918

Quant Time: Sep 28 09:56:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	225258	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	368685	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	360918	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199292	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	180525	56.10	ug/l	0.00
Spiked Amount			Recovery	=	112.20%	
35) Dibromofluoromethane	5.50	113	149057	47.70	ug/l	0.00
Spiked Amount			Recovery	=	95.40%	
50) Toluene-d8	8.72	98	539180	51.90	ug/l	0.00
Spiked Amount			Recovery	=	103.80%	
62) 4-Bromofluorobenzene	11.14	95	190952	51.03	ug/l	0.00
Spiked Amount			Recovery	=	102.06%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092718\
Data File : VX004804.D
Acq On : 28 Sep 2018 05:31
Operator : JC/MD
Sample : J5033-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPSI-TT-MW301I-20180918

Quant Time: Sep 28 09:56:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.01 min

Lab File: VX004804.D

Acq: 28 Sep 2018 05:31

Instrument :

MSVOA_X

ClientSampleId :


BPSI-TT-MW301I-20180918

Tot Ion:168 Resp: 225258

Ion Ratio Lower Upper

168 100

99 51.2 39.9 59.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 56.102 ug/l

RT: 6.07 min Scan# 807

Delta R.T. 0.01 min

Lab File: VX004804.D


Acq: 28 Sep 2018 05:31

Tgt Ion: 65 Resp: 180525

Ion Ratio Lower Upper

65 100

67 53.5 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

80000

60000


40000


20000

0

Time-->

5.90 6.00 6.10 6.20







#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX004804.D
 Acq: 28 Sep 2018 05:31

Instrument :
 MSVOA_X
 ClientSampleId :
 BPSI-TT-MW301I-20180918



Tot Ion:114 Resp: 368685
 Ion Ratio Lower Upper
 114 100
 63 21.2 0.0 39.0
 88 15.7 0.0 30.4



Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0




Time-->




#35
 Dibromofluoromethane
 Concen: 47.704 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. -0.00 min
 Lab File: VX004804.D
 Acq: 28 Sep 2018 05:31


Tgt Ion:113 Resp: 149057
 Ion Ratio Lower Upper
 113 100
 111 103.5 82.4 123.6
 192 22.9 19.4 29.2




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V



Time-->









#50
Toluene-d8
Concen: 51.905 ug/l
RT: 8.72 min Scan# 1241
Delta R.T. -0.00 min
Lab File: VX004804.D
Acq: 28 Sep 2018 05:31




Instrument :
MSVOA_X
ClientSampleId :
BPSI-TT-MW301I-20180918


Tot Ion: 98 Resp: 539180
Ion Ratio Lower Upper
98 100
100 62.9 52.9 79.3



#62
4-Bromofluorobenzene
Concen: 51.025 ug/l
RT: 11.14 min Scan# 1638
Delta R.T. -0.00 min
Lab File: VX004804.D
Acq: 28 Sep 2018 05:31

Tgt Ion: 95 Resp: 190952
Ion Ratio Lower Upper
95 100
174 92.3 0.0 185.2
176 88.1 0.0 178.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX004804.D
Acq: 28 Sep 2018 05:31

Instrument :
MSVOA_X
ClientSampleId :
BPSI-TT-MW301I-20180918

Tot Ion:117 Resp: 360918
Ion Ratio Lower Upper
117 100
82 55.1 47.8 71.6
119 32.1 29.3 43.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX004804.D
Acq: 28 Sep 2018 05:31

Tgt Ion:152 Resp: 199292
Ion Ratio Lower Upper
152 100
115 55.7 39.0 117.0
150 156.7 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

