


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092718\
Data File : VX004762.D
Acq On : 27 Sep 2018 09:06
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092118S.M
Title : SW846 8260
Last Update : Fri Sep 21 09:05:42 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	9954	PASS
75	95	30	60	48.5	28145	PASS
95	95	100	100	100.0	58072	PASS
96	95	5	9	6.9	4035	PASS
173	174	0.00	2	0.7	403	PASS
174	95	50	100	95.0	55194	PASS
175	174	5	9	7.0	3879	PASS
176	174	95	101	95.0	52437	PASS
177	176	5	9	6.4	3350	PASS