

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092719\
 Data File : VX012689.D
 Acq On : 27 Sep 2019 10:55
 Operator : JC/SP
 Sample : VX0927WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0927WBL01

Quant Time: Sep 28 04:12:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	163007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	285913	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	245528	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98246	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	116802	49.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.42%	
35) Dibromofluoromethane	5.48	113	86108	50.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.14%	
50) Toluene-d8	8.71	98	341154	52.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.06%	
62) 4-Bromofluorobenzene	11.13	95	120048	46.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092719\
Data File : VX012689.D
Acq On : 27 Sep 2019 10:55
Operator : JC/SP
Sample : VX0927WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0927WBL01

Quant Time: Sep 28 04:12:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

