

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092719\
 Data File : VX012690.D
 Acq On : 27 Sep 2019 11:18
 Operator : JC/SP
 Sample : K5049-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Sep 28 04:16:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	170090	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	286681	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	239821	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	86499	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	125687	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
35) Dibromofluoromethane	5.49	113	92320	54.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.14%	
50) Toluene-d8	8.71	98	358466	54.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.04%	
62) 4-Bromofluorobenzene	11.14	95	119155	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
Target Compounds						
16) Acetone	2.43	43	16931	12.905	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092719\
Data File : VX012690.D
Acq On : 27 Sep 2019 11:18
Operator : JC/SP
Sample : K5049-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Sep 28 04:16:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

