

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092723\
 Data File : VX037920.D
 Acq On : 27 Sep 2023 13:44
 Operator : JC/MD
 Sample : 04614-04ME
 Misc : 4.98g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1049ME

Quant Time: Sep 28 01:20:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	262583	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	434116	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	388335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	208242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	152398	44.174	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.340%
35) Dibromofluoromethane	5.385	113	126846	43.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.340%
50) Toluene-d8	8.647	98	516372	47.844	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.680%
62) 4-Bromofluorobenzene	11.086	95	185837	47.385	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.760%
Target Compounds						
16) Acetone	2.416	43	6445	3.883	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092723\
Data File : VX037920.D
Acq On : 27 Sep 2023 13:44
Operator : JC/MD
Sample : 04614-04ME
Misc : 4.98g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1049ME

Quant Time: Sep 28 01:20:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

