

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092723\
 Data File : VX037925.D
 Acq On : 27 Sep 2023 15:40
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 28 01:23:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	226505	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	383950	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	368513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193213	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139323	46.817	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.640%
35) Dibromofluoromethane	5.385	113	117711	45.819	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.640%
50) Toluene-d8	8.647	98	464800	48.693	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.380%
62) 4-Bromofluorobenzene	11.079	95	173779	50.100	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092723\
Data File : VX037925.D
Acq On : 27 Sep 2023 15:40
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 28 01:23:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

