

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092723\
 Data File : VX037935.D
 Acq On : 27 Sep 2023 20:22
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 28 06:11:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X092723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Sep 28 06:08:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	47080	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	261024	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	244075	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	100771	29.693	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.967%
60) 4-Bromofluorobenzene	11.079	95	111123	27.928	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.100%
63) Toluene-d8	8.647	98	317792	30.162	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.533%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092723\
Data File : VX037935.D
Acq On : 27 Sep 2023 20:22
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 28 06:11:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X092723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Sep 28 06:08:32 2023
Response via : Initial Calibration

