

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092818\
 Data File : VX004828.D
 Acq On : 28 Sep 2018 18:19
 Operator : JC/MD
 Sample : J5085-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-713

Quant Time: Sep 29 07:11:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	224874	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	368453	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	365506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	206948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	176762	55.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.06%	
35) Dibromofluoromethane	5.50	113	148485	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
50) Toluene-d8	8.72	98	535764	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
62) 4-Bromofluorobenzene	11.14	95	196965	52.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.34%	
Target Compounds						
16) Acetone	2.45	43	11699	6.238	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092818\
Data File : VX004828.D
Acq On : 28 Sep 2018 18:19
Operator : JC/MD
Sample : J5085-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-713

Quant Time: Sep 29 07:11:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

