

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092818\
 Data File : VX004833.D
 Acq On : 28 Sep 2018 20:32
 Operator : JC/MD
 Sample : J5148-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WS-1

Quant Time: Oct 01 13:06:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232760	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	377726	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	368074	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	202486	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	180867	54.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.80%	
35) Dibromofluoromethane	5.50	113	149918	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
50) Toluene-d8	8.71	98	546407	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
62) 4-Bromofluorobenzene	11.14	95	195994	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
Target Compounds						
3) Chloromethane	1.32	50	8922	2.317	ug/l	99
16) Acetone	2.45	43	146330	75.382	ug/l	96
25) 2-Butanone	4.71	43	7556	2.641	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092818\
Data File : VX004833.D
Acq On : 28 Sep 2018 20:32
Operator : JC/MD
Sample : J5148-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WS-1

Quant Time: Oct 01 13:06:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

