

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092818\
 Data File : VX004839.D
 Acq On : 29 Sep 2018 00:06
 Operator : JC/MD
 Sample : J5134-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20180924

Quant Time: Oct 01 12:40:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	234491	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	368041	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	361496	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	200274	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	182362	54.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.88%	
35) Dibromofluoromethane	5.50	113	153030	49.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.12%	
50) Toluene-d8	8.72	98	535822	51.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.34%	
62) 4-Bromofluorobenzene	11.14	95	191940	51.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.76%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092818\
Data File : VX004839.D
Acq On : 29 Sep 2018 00:06
Operator : JC/MD
Sample : J5134-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20180924

Quant Time: Oct 01 12:40:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

