

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092818\
 Data File : VX004834.D
 Acq On : 28 Sep 2018 20:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 29 07:25:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.682	0.803	-17.7	77	0.00
3 P	Chloromethane	0.827	0.861	-4.1	75	0.00
4 C	Vinyl Chloride	0.776	0.855	-10.2#	80	0.00
5 T	Bromomethane	0.460	0.382	17.0	62	0.00
6 T	Chloroethane	0.495	0.493	0.4	84	0.00
7 T	Trichlorofluoromethane	0.960	1.095	-14.1	80	0.00
8 T	Diethyl Ether	0.417	0.439	-5.3	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.547	0.608	-11.2	89	0.00
10 T	Methyl Iodide	0.732	0.456	37.7#	41#	0.00
11 T	Tert butyl alcohol	0.200	0.211	-5.5	81	0.00
12 CM	1,1-Dichloroethene	0.542	0.584	-7.7#	86	0.00
13 T	Acrolein	0.142	0.104	26.8#	58	0.00
14 T	Allyl chloride	1.182	1.314	-11.2	83	0.00
15 T	Acrylonitrile	0.440	0.474	-7.7	79	0.00
16 T	Acetone	0.417	0.413	1.0	73	0.00
17 T	Carbon Disulfide	1.628	1.694	-4.1	78	0.00
18 T	Methyl Acetate	1.056	1.226	-16.1	90	0.00
19 T	Methyl tert-butyl Ether	1.869	2.116	-13.2	89	0.00
20 T	Methylene Chloride	0.670	0.675	-0.7	76	0.00
21 T	trans-1,2-Dichloroethene	0.593	0.626	-5.6	78	0.00
22 T	Diisopropyl ether	1.991	2.236	-12.3	84	0.00
23 T	Vinyl Acetate	1.822	1.982	-8.8	78	0.00
24 P	1,1-Dichloroethane	1.187	1.262	-6.3	79	0.00
25 T	2-Butanone	0.615	0.638	-3.7	78	0.00
26 T	2,2-Dichloropropane	0.827	0.812	1.8	76	0.00
27 T	cis-1,2-Dichloroethene	0.663	0.692	-4.4	78	0.00
28 T	Bromochloromethane	0.559	0.562	-0.5	80	0.00
29 T	Tetrahydrofuran	0.382	0.419	-9.7	79	0.00
30 C	Chloroform	1.122	1.162	-3.6#	83	0.00
31 T	Cyclohexane	0.909	1.067	-17.4	81	0.00
32 T	1,1,1-Trichloroethane	0.916	0.980	-7.0	82	0.00
33 S	1,2-Dichloroethane-d4	0.714	0.723	-1.3	73	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.424	0.387	8.7	70	0.00
36 T	1,1-Dichloropropene	0.578	0.575	0.5	82	0.00
37 T	Ethyl Acetate	0.767	0.753	1.8	79	0.00
38 T	Carbon Tetrachloride	0.608	0.574	5.6	78	0.00
39 T	Methylcyclohexane	0.590	0.673	-14.1	94	0.00
40 TM	Benzene	1.810	1.742	3.8	81	0.00
41 T	Methacrylonitrile	0.426	0.432	-1.4	79	0.00
42 TM	1,2-Dichloroethane	0.640	0.618	3.4	80	0.00
43 T	Isopropyl Acetate	1.108	1.133	-2.3	77	0.00
44 TM	Trichloroethene	0.445	0.455	-2.2	91	0.00
45 C	1,2-Dichloropropane	0.444	0.466	-5.0#	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092818\
 Data File : VX004834.D
 Acq On : 28 Sep 2018 20:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 29 07:25:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.279	0.298	-6.8	93	0.00
47 T	Bromodichloromethane	0.522	0.568	-8.8	92	0.00
48 T	Methyl methacrylate	0.503	0.592	-17.7	93	0.00
49 T	1,4-Dioxane	0.012	0.013	-8.3	80	0.00
50 S	Toluene-d8	1.409	1.477	-4.8	90	0.00
51 T	4-Methyl-2-Pentanone	0.697	0.806	-15.6	93	0.00
52 CM	Toluene	0.995	1.086	-9.1#	88	0.00
53 T	t-1,3-Dichloropropene	0.590	0.639	-8.3	82	0.00
54 T	cis-1,3-Dichloropropene	0.618	0.703	-13.8	93	0.00
55 T	1,1,2-Trichloroethane	0.436	0.445	-2.1	82	0.00
56 T	Ethyl methacrylate	0.628	0.716	-14.0	87	0.00
57 T	1,3-Dichloropropane	0.725	0.760	-4.8	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.331	0.389	-17.5	89	0.00
59 T	2-Hexanone	0.581	0.634	-9.1	82	0.00
60 T	Dibromochloromethane	0.436	0.459	-5.3	78	0.00
61 T	1,2-Dibromoethane	0.467	0.465	0.4	76	0.00
62 S	4-Bromofluorobenzene	0.508	0.549	-8.1	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.492	0.456	7.3	78	0.00
65 PM	Chlorobenzene	1.282	1.262	1.6	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.451	0.454	-0.7	89	0.00
67 C	Ethyl Benzene	2.036	2.181	-7.1#	93	0.00
68 T	m/p-Xylenes	0.795	0.842	-5.9	88	0.00
69 T	o-Xylene	0.731	0.812	-11.1	94	0.00
70 T	Styrene	1.227	1.390	-13.3	94	0.00
71 P	Bromoform	0.386	0.394	-2.1	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.152	3.533	-12.1	95	0.00
74 T	N-amyl acetate	1.531	1.761	-15.0	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.283	1.253	2.3	94	0.00
76 T	1,2,3-Trichloropropane	1.121	1.147	-2.3	94	0.00
77 T	Bromobenzene	0.915	0.944	-3.2	94	0.00
78 T	n-propylbenzene	3.616	4.079	-12.8	95	0.00
79 T	2-Chlorotoluene	2.254	2.426	-7.6	94	0.00
80 T	1,3,5-Trimethylbenzene	2.664	3.020	-13.4	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.367	0.359	2.2	88	0.00
82 T	4-Chlorotoluene	2.653	2.879	-8.5	94	0.00
83 T	tert-Butylbenzene	2.595	2.949	-13.6	97	0.00
84 T	1,2,4-Trimethylbenzene	2.728	3.108	-13.9	94	0.00
85 T	sec-Butylbenzene	3.161	3.568	-12.9	96	0.00
86 T	p-Isopropyltoluene	2.863	3.226	-12.7	96	0.00
87 T	1,3-Dichlorobenzene	1.709	1.736	-1.6	93	0.00
88 T	1,4-Dichlorobenzene	1.765	1.746	1.1	92	0.00
89 T	n-Butylbenzene	2.601	2.881	-10.8	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092818\
Data File : VX004834.D
Acq On : 28 Sep 2018 20:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 29 07:25:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.477	0.512	-7.3	102	0.00
91 T	1,2-Dichlorobenzene	1.786	1.766	1.1	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.282	0.301	-6.7	100	0.00
93 T	1,2,4-Trichlorobenzene	1.253	1.313	-4.8	94	0.00
94 T	Hexachlorobutadiene	0.650	0.620	4.6	92	0.00
95 T	Naphthalene	3.727	4.244	-13.9	94	0.00
96 T	1,2,3-Trichlorobenzene	1.294	1.350	-4.3	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6