

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092818\
 Data File : VX004842.D
 Acq On : 29 Sep 2018 01:26
 Operator : JC/MD
 Sample : J5134-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW504-2125

Quant Time: Oct 01 12:42:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	238680	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	386014	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	377648	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	212144	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	190506	55.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.74%	
35) Dibromofluoromethane	5.50	113	162446	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
50) Toluene-d8	8.72	98	561995	51.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.34%	
62) 4-Bromofluorobenzene	11.14	95	202986	51.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.62%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092818\
Data File : VX004842.D
Acq On : 29 Sep 2018 01:26
Operator : JC/MD
Sample : J5134-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW504-2125

Quant Time: Oct 01 12:42:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.01 min

Lab File: VX004842.D

Acq: 29 Sep 2018 01:26

Instrument :

MSVOA_X

ClientSampleId :


SA-TW504-2125

Tot Ion:168 Resp: 238680

Ion Ratio Lower Upper




168 100

99 49.9 39.9 59.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 55.874 ug/l

RT: 6.07 min Scan# 807

Delta R.T. 0.01 min

Lab File: VX004842.D


Acq: 29 Sep 2018 01:26

Tgt Ion: 65 Resp: 190506

Ion Ratio Lower Upper



65 100


67 51.8 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX004842.D
 Acq: 29 Sep 2018 01:26

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW504-2125

Tot Ion:114 Resp: 386014
 Ion Ratio Lower Upper
 114 100
 63 22.3 0.0 39.0
 88 15.8 0.0 30.4




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 49.655 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. -0.00 min
 Lab File: VX004842.D
 Acq: 29 Sep 2018 01:26

Tgt Ion:113 Resp: 162446
 Ion Ratio Lower Upper
 113 100
 111 100.4 82.4 123.6
 192 22.9 19.4 29.2



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 51.672 ug/l
RT: 8.72 min Scan# 1241
Delta R.T. -0.00 min
Lab File: VX004842.D
Acq: 29 Sep 2018 01:26




Instrument :
MSVOA_X
ClientSampleId :
SA-TW504-2125


Tot Ion: 98 Resp: 561995
Ion Ratio Lower Upper
98 100
100 63.5 52.9 79.3



#62
4-Bromofluorobenzene
Concen: 51.806 ug/l
RT: 11.14 min Scan# 1638
Delta R.T. -0.00 min
Lab File: VX004842.D
Acq: 29 Sep 2018 01:26

Tgt Ion: 95 Resp: 202986
Ion Ratio Lower Upper
95 100
174 91.2 0.0 185.2
176 88.2 0.0 178.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX004842.D
Acq: 29 Sep 2018 01:26

Instrument :
MSVOA_X
ClientSampleId :
SA-TW504-2125

Tot Ion:117 Resp: 377648
Ion Ratio Lower Upper
117 100
82 54.7 47.8 71.6
119 32.4 29.3 43.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX004842.D
Acq: 29 Sep 2018 01:26

Tgt Ion:152 Resp: 212144
Ion Ratio Lower Upper
152 100
115 55.7 39.0 117.0
150 158.1 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

