

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VOX092820\
 Data File : VX018621.D
 Acq On : 28 Sep 2020 17:42
 Operator : JC/SP
 Sample : L4135-01
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 29 03:00:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 29 02:30:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	279585	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	271254	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	145295	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	62625	4.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.60%
7) Chloroethane-d5	1.70	69	62353	5.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.00%
11) 1,1-Dichloroethene-d2	2.34	63	110472	3.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.40%
20) 2-Butanone-d5	4.56	46	162293	43.50	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.00%
24) Chloroform-d	5.14	84	148171	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.40%
26) 1,2-Dichloroethane-d4	6.04	65	80712	4.00	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.00%
32) Benzene-d6	6.06	84	277482	4.43	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
36) 1,2-Dichloropropane-d6	7.37	67	84795	4.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.80%
41) Toluene-d8	8.70	98	267417	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.80%
45) trans-1,3-Dichloropropene-	8.99	79	34695	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
48) 2-Hexanone-d5	9.43	63	137727	45.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.64%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	71097	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
65) 1,2-Dichlorobenzene-d4	12.36	152	109286	4.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.00%

Target Compounds

					Ovalue
5) Vinyl chloride	1.39	62	28992	1.915	ug/L 100
8) Chloroethane	1.72	64	158566	19.048	ug/L 97
13) Acetone	2.46	43	50640	24.524	ug/L 94
14) Carbon disulfide	2.60	76	195390	4.708	ug/L 97
16) Methylene chloride	2.83	84	5474	0.294	ug/L 93
18) trans-1,2-Dichloroethene	3.14	96	24734	1.813	ug/L 89
19) 1,1-Dichloroethane	3.67	63	27686	1.132	ug/L # 91
22) cis-1,2-Dichloroethene	4.57	96	32870	2.309	ug/L # 91
30) Cyclohexane	5.55	56	287582	13.993	ug/L 98
33) Benzene	6.11	78	32604779m	623.431	ug/L 97
35) Methylcyclohexane	7.44	83	264749	12.577	ug/L 97
40) 4-Methyl-2-pentanone	8.62	43	7565m	0.978	ug/L 75
42) Toluene	8.76	91	19986012	353.794	ug/L # 75
53) Chlorobenzene	10.12	112	20019967m	536.478	ug/L 97
54) Ethylbenzene	10.24	91	2317073	37.655	ug/L 96
55) m,p-xylene	10.34	106	511859	21.755	ug/L 87
56) o-xylene	10.68	106	897985	40.392	ug/L 87

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092820\
Data File : VX018621.D
Acq On : 28 Sep 2020 17:42
Operator : JC/SP
Sample : L4135-01
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Sep 29 03:00:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 29 02:30:52 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
58) Isopropylbenzene	11.00	105	3517065	59.003	ug/L	100
63) 1,3-Dichlorobenzene	12.01	146	295852	9.367	ug/L	98
64) 1,4-Dichlorobenzene	12.09	146	1529297	47.932	ug/L	96
66) 1,2-Dichlorobenzene	12.38	146	636737	21.638	ug/L	97
69) 1,2,4-trichlorobenzene	13.63	180	3741	0.192	ug/L #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

