

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092820\
 Data File : VX018640.D
 Acq On : 29 Sep 2020 01:22
 Operator : JC/SP
 Sample : L4135-20
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BG498

Quant Time: Sep 29 06:22:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 29 02:30:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	308508	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	295070	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	137450	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	74349	4.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.20%
7) Chloroethane-d5	1.70	69	63930	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.00%
11) 1,1-Dichloroethene-d2	2.34	63	129229	3.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.40%
20) 2-Butanone-d5	4.55	46	223784	54.36	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.72%
24) Chloroform-d	5.14	84	174100	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
26) 1,2-Dichloroethane-d4	6.03	65	97822	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.80%
32) Benzene-d6	6.05	84	338976	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
36) 1,2-Dichloropropane-d6	7.37	67	104998	5.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.20%
41) Toluene-d8	8.70	98	311455	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
45) trans-1,3-Dichloropropene-	8.99	79	42596	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
48) 2-Hexanone-d5	9.43	63	179398	54.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.54%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	77653	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
65) 1,2-Dichlorobenzene-d4	12.36	152	116107	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chloromethane	1.31	50	2842	0.178	ug/L	97
33) Benzene	6.12	78	17157	0.302	ug/L	100
42) Toluene	8.76	91	27218	0.443	ug/L	95
53) Chlorobenzene	10.12	112	106269	2.618	ug/L	96
55) m,p-xylene	10.35	106	3130	0.122	ug/L	73
64) 1,4-Dichlorobenzene	12.08	146	13289	0.440	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092820\
Data File : VX018640.D
Acq On : 29 Sep 2020 01:22
Operator : JC/SP
Sample : L4135-20
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG498

Quant Time: Sep 29 06:22:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 29 02:30:52 2020
Response via : Initial Calibration

