

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
 Data File : VX031717.D
 Acq On : 28 Sep 2022 11:14
 Operator : JC/MD
 Sample : VX0928MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0928MBL01

Quant Time: Sep 29 04:29:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	126899	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.757	114	206225	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107367	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	131055	48.659	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.320%
35) Dibromofluoromethane	5.379	113	110577	45.001	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.000%
50) Toluene-d8	8.647	98	414009	46.474	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.940%
62) 4-Bromofluorobenzene	11.079	95	141130	43.224	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
Data File : VX031717.D
Acq On : 28 Sep 2022 11:14
Operator : JC/MD
Sample : VX0928MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0928MBL01

Quant Time: Sep 29 04:29:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

