

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
 Data File : VX031718.D
 Acq On : 28 Sep 2022 11:37
 Operator : JC/MD
 Sample : VX0928WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0928WBL01

Quant Time: Sep 29 04:29:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	122773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198345	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199453	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104778	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	126362	48.494	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.980%
35) Dibromofluoromethane	5.385	113	102338	43.303	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.600%
50) Toluene-d8	8.647	98	393636	45.942	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.880%
62) 4-Bromofluorobenzene	11.079	95	132773	42.280	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
Data File : VX031718.D
Acq On : 28 Sep 2022 11:37
Operator : JC/MD
Sample : VX0928WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0928WBL01

Quant Time: Sep 29 04:29:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

