

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092822\  
 Data File : VX031723.D  
 Acq On : 28 Sep 2022 13:43  
 Operator : JC/MD  
 Sample : N4848-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BP-VPB-RW5-GW-628-630

Quant Time: Sep 29 04:30:55 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092622W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Sep 27 01:44:18 2022  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|------------------------------|--------|-------|----------|----------|------------|----------|
| -----                        |        |       |          |          |            |          |
| Internal Standards           |        |       |          |          |            |          |
| 1) Pentafluorobenzene        | 5.550  | 168   | 119535   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763  | 114   | 190601   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5         | 10.055 | 117   | 197238   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024 | 152   | 103920   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds  |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4    | 5.958  | 65    | 127084   | 50.092   | ug/l       | 0.00     |
| Spiked Amount                | 50.000 | Range | 74 - 125 | Recovery | = 100.180% |          |
| 35) Dibromofluoromethane     | 5.385  | 113   | 103522   | 45.583   | ug/l       | 0.00     |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | = 91.160%  |          |
| 50) Toluene-d8               | 8.647  | 98    | 403645   | 49.025   | ug/l       | 0.00     |
| Spiked Amount                | 50.000 | Range | 86 - 113 | Recovery | = 98.040%  |          |
| 62) 4-Bromofluorobenzene     | 11.079 | 95    | 136530   | 45.243   | ug/l       | 0.00     |
| Spiked Amount                | 50.000 | Range | 83 - 123 | Recovery | = 90.480%  |          |
| Target Compounds             |        |       |          |          |            | Qvalue   |
| 9) 1,1,2-Trichlorotrifluo... | 2.331  | 101   | 2408     | 1.515    | ug/l       | 92       |
| 30) Chloroform               | 5.093  | 83    | 1518     | 0.398    | ug/l #     | 79       |
| 44) Trichloroethene          | 7.123  | 130   | 35062    | 15.497   | ug/l       | 92       |
| -----                        |        |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092822\  
Data File : VX031723.D  
Acq On : 28 Sep 2022 13:43  
Operator : JC/MD  
Sample : N4848-03  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BP-VPB-RW5-GW-628-630

Quant Time: Sep 29 04:30:55 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092622W.M  
Quant Title : SW846 8260  
QLast Update : Tue Sep 27 01:44:18 2022  
Response via : Initial Calibration

