

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
 Data File : VX031726.D
 Acq On : 28 Sep 2022 14:54
 Operator : JC/MD
 Sample : N4848-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-EB-20220926

Quant Time: Sep 29 04:31:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	116079	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	182797	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	126346	51.284	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.560%
35) Dibromofluoromethane	5.385	113	102551	47.084	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.160%
50) Toluene-d8	8.647	98	392731	49.736	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.480%
62) 4-Bromofluorobenzene	11.079	95	134503	46.474	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%
Target Compounds						
30) Chloroform	5.099	83	2689	0.726	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
Data File : VX031726.D
Acq On : 28 Sep 2022 14:54
Operator : JC/MD
Sample : N4848-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220926

Quant Time: Sep 29 04:31:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

