

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
 Data File : VX031727.D
 Acq On : 28 Sep 2022 15:17
 Operator : JC/MD
 Sample : N4846-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20220923

Quant Time: Sep 29 04:31:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	188918	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105700	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	124176	50.458	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.920%
35) Dibromofluoromethane	5.385	113	100837	44.797	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.600%
50) Toluene-d8	8.647	98	386421	47.351	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.700%
62) 4-Bromofluorobenzene	11.079	95	131267	43.886	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
Data File : VX031727.D
Acq On : 28 Sep 2022 15:17
Operator : JC/MD
Sample : N4846-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20220923

Quant Time: Sep 29 04:31:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

