

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
 Data File : VX031735.D
 Acq On : 28 Sep 2022 18:24
 Operator : JC/MD
 Sample : N4793-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 29 04:33:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	111482	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	185618	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191502	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	124189	52.487	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.980%	
35) Dibromofluoromethane	5.385	113	102109	46.168	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.340%	
50) Toluene-d8	8.647	98	387997	48.389	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.780%	
62) 4-Bromofluorobenzene	11.079	95	129461	44.052	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.100%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
Data File : VX031735.D
Acq On : 28 Sep 2022 18:24
Operator : JC/MD
Sample : N4793-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 29 04:33:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

