

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
 Data File : VX031752.D
 Acq On : 29 Sep 2022 02:10
 Operator : JC/MD
 Sample : N4863-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Sep 29 04:37:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	125287	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205168	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	135357	50.903	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.800%
35) Dibromofluoromethane	5.385	113	113596	46.468	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.940%
50) Toluene-d8	8.647	98	434169	48.988	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.980%
62) 4-Bromofluorobenzene	11.079	95	146062	44.965	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.920%
Target Compounds						
19) Methyl tert-butyl Ether	3.111	73	416526	70.427	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
Data File : VX031752.D
Acq On : 29 Sep 2022 02:10
Operator : JC/MD
Sample : N4863-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Sep 29 04:37:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

