

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
 Data File : VX031757.D
 Acq On : 29 Sep 2022 04:06
 Operator : JC/MD
 Sample : N4812-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-3

Quant Time: Sep 29 06:25:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	123506	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	207550	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	143120	54.599	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.200%	
35) Dibromofluoromethane	5.385	113	119842	48.460	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.920%	
50) Toluene-d8	8.647	98	478693	53.392	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.780%	
62) 4-Bromofluorobenzene	11.079	95	169113	51.464	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.920%	
Target Compounds						
						Qvalue
20) Methylene Chloride	2.788	84	6379	2.231	ug/l #	84
37) Ethyl Acetate	4.721	43	11404	3.981	ug/l #	93
43) Isopropyl Acetate	6.342	43	54123	12.004	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092822\
Data File : VX031757.D
Acq On : 29 Sep 2022 04:06
Operator : JC/MD
Sample : N4812-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-3

Quant Time: Sep 29 06:25:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

