

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092823\
 Data File : VX037944.D
 Acq On : 28 Sep 2023 12:10
 Operator : JC/MD
 Sample : 04346-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 29 04:24:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X092723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Sep 28 06:08:32 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Bromochloromethane		4.897	128	44114	30.000	ug/l	0.00
28) 1,4-Difluorobenzene		6.763	114	251016	30.000	ug/l	0.00
57) Chlorobenzene-d5		10.055	117	243641	30.000	ug/l	0.00
System Monitoring Compounds							
27) 1,2-Dichloroethane-d4		5.952	65	96612	30.382	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.267%	
60) 4-Bromofluorobenzene		11.079	95	115807	29.157	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	97.200%	
63) Toluene-d8		8.647	98	313099	29.769	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.233%	
Target Compounds							Qvalue
4) Vinyl Chloride		1.691	62	5641	2.035	ug/l	# 1
5) Bromomethane		1.617	94	739	0.346	ug/l	83
6) Chloroethane		1.691	64	157964	84.869	ug/l	95
7) Trichlorofluoromethane		1.892	101	384747	84.071	ug/l	100
9) 1,1,2-Trichlorotrifluo...		2.318	101	1092	0.468	ug/l	39
10) 1,1-Dichloroethene		2.318	96	340653	149.148	ug/l	98
15) Acetone		2.379	58	18343	49.475	ug/l	86
16) Carbon Disulfide		2.514	76	1134	0.203	ug/l	99
18) Methylene Chloride		2.788	84	246539	90.062	ug/l	95
19) trans-1,2-Dichloroethene		3.093	96	65687	25.517	ug/l	96
21) 1,1-Dichloroethane		3.611	63	615127	136.383	ug/l	99
22) cis-1,2-Dichloroethene		4.489	96	410225	134.213	ug/l	87
23) tert-Butyl Alcohol		3.093	59	3056	5.714	ug/l	# 1
25) Chloroform		5.092	83	445530	89.821	ug/l	100
30) 2-Butanone		4.550	43	309700	177.847	ug/l	96
32) 1,1,1-Trichloroethane		5.385	97	337426	77.606	ug/l	98
33) Carbon Tetrachloride		5.678	117	416713	111.961	ug/l	99
38) Methylcyclohexane		7.433	83	2269	0.497	ug/l	# 25
39) 1,2-Dichloropropane		7.427	63	315880	113.908	ug/l	96
40) Dibromomethane		7.580	93	39777	19.439	ug/l	98
41) Bromodichloromethane		7.824	83	1148	0.307	ug/l	# 84
45) 1,4-Dioxane		7.586	88	187	2.849	ug/l	# 1
48) t-1,3-Dichloropropene		8.976	75	375823	92.603	ug/l	98
49) cis-1,3-Dichloropropene		8.366	75	376379	86.356	ug/l	94
50) 1,1,2-Trichloroethane		9.153	97	399316	136.557	ug/l	97
51) Ethyl methacrylate		8.653	69	896	0.208	ug/l	85
53) Dibromochloromethane		9.518	129	299318	99.319	ug/l	100
54) 1,2-Dibromoethane		9.610	107	335372	105.690	ug/l	99
55) 2-Chloroethyl vinyl ether		8.269	63	856	0.408	ug/l	# 74
56) Bromoform		10.799	173	206317	91.022	ug/l	98
61) Tetrachloroethene		9.275	164	293898	113.351	ug/l	97
62) Toluene		8.720	91	292518	23.145	ug/l	100
64) Chlorobenzene		10.079	112	389185	49.364	ug/l	100
66) Ethyl Benzene		10.195	91	842298	60.387	ug/l	97
67) m/p-Xylenes		10.299	106	210182	38.251	ug/l	98
68) o-Xylene		10.640	106	174209	32.740	ug/l	96
69) Styrene		10.640	104	8973	1.005	ug/l	# 1
71) 1,1,2,2-Tetrachloroethane		11.238	83	6383	1.371	ug/l	# 95
72) 1,2,3-Trichloropropane		11.238	75	221479	54.197	ug/l	90

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092823\
 Data File : VX037944.D
 Acq On : 28 Sep 2023 12:10
 Operator : JC/MD
 Sample : 04346-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 29 04:24:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X092723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Sep 28 06:08:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


76) 1,3,5-Trimethylbenzene	11.451	105	1152559	94.645	ug/l	100
78) 4-Chlorotoluene	11.451	91	123924	10.696	ug/l	45
80) 1,2,4-Trimethylbenzene	11.756	105	6740	0.557	ug/l	92
83) 1,3-Dichlorobenzene	11.969	146	650119	93.937	ug/l	99
84) 1,4-Dichlorobenzene	12.042	146	693700	98.918	ug/l	99
87) 1,2-Dichlorobenzene	12.335	146	129032	18.923	ug/l	97
88) 1,2-Dibromo-3-Chloropr...	12.939	75	146000	157.495	ug/l	84
89) 1,2,4-Trichlorobenzene	13.591	180	1181	0.255	ug/l	97
91) Naphthalene	13.774	128	2059590	146.377	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092823\
Data File : VX037944.D
Acq On : 28 Sep 2023 12:10
Operator : JC/MD
Sample : 04346-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Sep 29 04:24:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X092723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Sep 28 06:08:32 2023
Response via : Initial Calibration








#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 4.897 min Scan# 61
Delta R.T. -0.007 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:128 Resp: 44114
Ion Ratio Lower Upper
128 100
49 142.8 0.0 406.3
130 129.3 0.0 328.0



#4
Vinyl Chloride
Concen: 2.035 ug/l
RT: 1.691 min Scan# 99
Delta R.T. 0.317 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion: 62 Resp: 5641
Ion Ratio Lower Upper
62 100
64 2896.1 25.3 37.9#





#5

Bromomethane

Concen: 0.346 ug/l

RT: 1.617 min Scan# 81

Delta R.T. 0.006 min


Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 94 Resp: 739

Ion Ratio Lower Upper

94 100

96 113.9 77.4 116.2



#6

Chloroethane

Concen: 84.869 ug/l

RT: 1.691 min Scan# 99

Delta R.T. -0.000 min

Lab File: VX037944.D




Acq: 28 Sep 2023 12:10


Tgt Ion: 64 Resp: 157964

Ion Ratio Lower Upper

64 100


66 31.6 23.1 34.7








#7
 Trichlorofluoromethane
 Concen: 84.071 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX037944.D
 Acq: 28 Sep 2023 12:10


Instrument :
 MSVOA_X
 ClientSampleId :





Tgt Ion:101 Resp: 384747
 Ion Ratio Lower Upper
 101 100
 103 65.2 52.2 78.2




#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.468 ug/l
 RT: 2.318 min Scan# 202
 Delta R.T. -0.012 min
 Lab File: VX037944.D
 Acq: 28 Sep 2023 12:10



Tgt Ion:101 Resp: 1092
 Ion Ratio Lower Upper
 101 100
 85 18.8 0.0 91.0
 151 14.7 0.0 155.4





#10

1,1-Dichloroethene

Concen: 149.148 ug/l

RT: 2.318 min Scan# 20

Delta R.T. -0.006 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 340653

Ion Ratio Lower Upper

96 100

61 155.9 127.2 190.8

98 64.0 52.7 79.1



#15

Acetone

Concen: 49.475 ug/l

RT: 2.379 min Scan# 212

Delta R.T. 0.006 min

Lab File: VX037944.D




Acq: 28 Sep 2023 12:10


Tgt Ion: 58 Resp: 18343

Ion Ratio Lower Upper

58 100

43 296.3 261.0 391.4









#16
Carbon Disulfide
Concen: 0.203 ug/l
RT: 2.514 min Scan# 211
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 1134
Ion Ratio Lower Upper
76 100
78 9.5 8.0 12.0



#18
Methylene Chloride
Concen: 90.062 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion: 84 Resp: 246539
Ion Ratio Lower Upper
84 100
49 113.1 95.4 143.0
51 35.7 28.3 42.5
86 65.6 49.0 73.6





#19

trans-1,2-Dichloroethene

Concen: 25.517 ug/l

RT: 3.093 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 65687

Ion Ratio Lower Upper

96 100

61 133.9 108.7 163.1

98 67.9 49.1 73.7



#21

1,1-Dichloroethane

Concen: 136.383 ug/l

RT: 3.611 min Scan# 414

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10




Tgt Ion: 63 Resp: 615127


Ion Ratio Lower Upper

63 100

98 7.0 3.2 9.6

100 4.5 2.3 6.8





#22

cis-1,2-Dichloroethene

Concen: 134.213 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 410225

Ion Ratio Lower Upper

96 100

61 119.0 114.0 171.0

98 64.1 51.6 77.4



#23

tert-Butyl Alcohol

Concen: 5.714 ug/l

RT: 3.093 min Scan# 329

Delta R.T. 0.140 min

Lab File: VX037944.D




Acq: 28 Sep 2023 12:10


Tgt Ion: 59 Resp: 3056

Ion Ratio Lower Upper

59 100

52 2.7 142.6 213.8#









#25
Chloroform
Concen: 89.821 ug/l
RT: 5.092 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 445530
Ion Ratio Lower Upper
83 100
85 64.7 51.5 77.3



#27
1,2-Dichloroethane-d4
Concen: 30.382 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.006 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion: 65 Resp: 96612
Ion Ratio Lower Upper
65 100
67 52.9 42.1 63.1





#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 6.763 min Scan# 91

Delta R.T. -0.000 min


Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:114 Resp: 251016

Ion Ratio Lower Upper

114 100

63 18.9 16.6 24.8

88 15.1 12.9 19.3



#30

2-Butanone

Concen: 177.847 ug/l

RT: 4.550 min Scan# 568

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10




Tgt Ion: 43 Resp: 309700


Ion Ratio Lower Upper

43 100

57 8.6 6.4 9.6

72 29.1 21.6 32.4





#32

1,1,1-Trichloroethane

Concen: 77.606 ug/l

RT: 5.385 min Scan# 705

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 97 Resp: 337426

Ion Ratio Lower Upper

97 100

99 64.8 51.5 77.3

61 42.3 35.8 53.8



#33

Carbon Tetrachloride

Concen: 111.961 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10




Tgt Ion: 117 Resp: 416713


Ion Ratio Lower Upper

117 100

119 95.6 76.6 115.0

121 31.1 26.5 39.7





#38

Methylcyclohexane

Concen: 0.497 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. 0.049 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 2269

Ion Ratio Lower Upper

83 100

55 0.0 38.1 114.5#

98 85.3 24.4 73.4#



#39

1,2-Dichloropropane

Concen: 113.908 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX037944.D




Acq: 28 Sep 2023 12:10


Tgt Ion: 63 Resp: 315880

Ion Ratio Lower Upper

63 100


65 31.0 26.4 39.6








#40
Dibromomethane
Concen: 19.439 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 93 Resp: 39777
Ion Ratio Lower Upper
93 100
95 82.7 0.0 166.8
174 98.9 0.0 191.6




#41
Bromodichloromethane
Concen: 0.307 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10




Tgt Ion: 83 Resp: 1148
Ion Ratio Lower Upper
83 100
85 52.7 51.5 77.3
127 0.0 6.4 9.6#








#45
1,4-Dioxane
Concen: 2.849 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.073 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 88 Resp: 187
Ion Ratio Lower Upper
88 100
43 23.5 297.5 446.3#
58 0.0 1.0 1.4#



#48
t-1,3-Dichloropropene
Concen: 92.603 ug/l
RT: 8.976 min Scan# 1294
Delta R.T. -0.006 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion: 75 Resp: 375823
Ion Ratio Lower Upper
75 100
77 32.1 26.8 40.2





#49

cis-1,3-Dichloropropene

Concen: 86.356 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 75 Resp: 376379

Ion Ratio Lower Upper

75 100

77 31.6 25.4 38.2

39 44.2 40.2 60.4



#50

1,1,2-Trichloroethane

Concen: 136.557 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Tgt Ion: 97 Resp: 399316




Ion Ratio Lower Upper


97 100

83 85.3 70.0 105.0

85 54.1 46.1 69.1

99 63.0 52.5 78.7





#51

Ethyl methacrylate

Concen: 0.208 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.463 min


Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 69 Resp: 896

Ion Ratio Lower Upper

69 100

41 74.4 32.0 96.0

39 28.6 19.4 58.4



#53

Dibromochloromethane

Concen: 99.319 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX037944.D




Acq: 28 Sep 2023 12:10


Tgt Ion: 129 Resp: 299318

Ion Ratio Lower Upper

129 100

127 77.3 38.6 115.8









#54
1,2-Dibromoethane
Concen: 105.690 ug/l
RT: 9.610 min Scan# 1178
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 335372
Ion Ratio Lower Upper
107 100
109 94.8 74.7 112.1



#55
2-Chloroethyl vinyl ether
Concen: 0.408 ug/l
RT: 8.269 min Scan# 1178
Delta R.T. 0.024 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion: 63 Resp: 856
Ion Ratio Lower Upper
63 100
65 12.5 25.9 38.9#
106 19.4 21.8 32.8#





#56

Bromoform

Concen: 91.022 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min


Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:173 Resp: 206317

Ion Ratio Lower Upper

173 100

175 49.1 38.1 57.1

254 9.2 6.7 10.1



#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10




Tgt Ion:117 Resp: 243641


Ion Ratio Lower Upper

117 100

82 53.7 44.7 67.1

119 31.9 25.6 38.4









#60
4-Bromofluorobenzene
Concen: 29.157 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 95 Resp: 115807
Ion Ratio Lower Upper
95 100
174 83.0 60.5 90.7
176 80.6 58.2 87.4



#61
Tetrachloroethene
Concen: 113.351 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion: 164 Resp: 293898
Ion Ratio Lower Upper
164 100
166 125.8 104.5 156.7
129 96.8 78.0 117.0
131 93.8 77.5 116.3









#62
Toluene
Concen: 23.145 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 292518
Ion Ratio Lower Upper
91 100
92 56.3 44.9 67.3



#63
Toluene-d8
Concen: 29.769 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10


Tgt Ion: 98 Resp: 313099
Ion Ratio Lower Upper
98 100
100 64.1 51.6 77.4








#64
Chlorobenzene
Concen: 49.364 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:112 Resp: 389185
Ion Ratio Lower Upper
112 100
114 32.1 25.8 38.6



#66
Ethyl Benzene
Concen: 60.387 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion: 91 Resp: 842298
Ion Ratio Lower Upper
91 100
106 31.9 24.2 36.4









#67
m/p-Xylenes
Concen: 38.251 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 210182
Ion Ratio Lower Upper
106 100
91 203.1 160.5 240.7



#68
o-Xylene
Concen: 32.740 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion:106 Resp: 174209
Ion Ratio Lower Upper
106 100
91 216.7 105.4 316.1





#69

Styrene

Concen: 1.005 ug/l

RT: 10.640 min Scan# 11

Delta R.T. -0.012 min


Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:104 Resp: 8973

Ion Ratio Lower Upper

104 100

78 292.5 41.0 61.4#

103 272.3 45.5 68.3#



#71

1,1,2,2-Tetrachloroethane

Concen: 1.371 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.024 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10




Tgt Ion: 83 Resp: 6383


Ion Ratio Lower Upper

83 100

131 0.6 5.3 15.8#

85 65.8 32.4 97.0









#72
1,2,3-Trichloropropane
Concen: 54.197 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 221479
Ion Ratio Lower Upper
75 100
77 32.5 0.0 77.2



#76
1,3,5-Trimethylbenzene
Concen: 94.645 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion:105 Resp: 1152559
Ion Ratio Lower Upper
105 100
120 48.4 0.0 96.4









#78
4-Chlorotoluene
Concen: 10.696 ug/l
RT: 11.451 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 123924
Ion Ratio Lower Upper
91 100
126 0.2 0.0 59.4



#80
1,2,4-Trimethylbenzene
Concen: 0.557 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VX037944.D
Acq: 28 Sep 2023 12:10

Tgt Ion: 105 Resp: 6740
Ion Ratio Lower Upper
105 100
120 50.3 0.0 90.2





#83

1,3-Dichlorobenzene

Concen: 93.937 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min


Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:146 Resp: 650119

Ion Ratio Lower Upper

146 100

111 40.0 20.4 61.2

148 63.9 31.6 94.8



#84

1,4-Dichlorobenzene

Concen: 98.918 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10




Tgt Ion:146 Resp: 693700


Ion Ratio Lower Upper

146 100

111 39.0 20.2 60.6


148 63.9 32.1 96.3








#87
 1,2-Dichlorobenzene
 Concen: 18.923 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. -0.000 min
 Lab File: VX037944.D
 Acq: 28 Sep 2023 12:10


Instrument :
 MSVOA_X
 ClientSampleId :





Tgt Ion:146 Resp: 129032
 Ion Ratio Lower Upper
 146 100
 111 41.4 21.4 64.3
 148 63.4 33.2 99.6




#88
 1,2-Dibromo-3-Chloropropane
 Concen: 157.495 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. -0.006 min
 Lab File: VX037944.D
 Acq: 28 Sep 2023 12:10



Tgt Ion: 75 Resp: 146000
 Ion Ratio Lower Upper
 75 100
 155 91.9 62.7 94.1
 157 118.3 81.5 122.3





#89

1,2,4-Trichlorobenzene

Concen: 0.255 ug/l

RT: 13.591 min Scan# 2050

Delta R.T. 0.006 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 1181

Ion Ratio Lower Upper

180 100

182 93.1 75.6 113.4

145 35.2 25.4 38.2



#91

Naphthalene

Concen: 146.377 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX037944.D

Acq: 28 Sep 2023 12:10

Tgt Ion:128 Resp: 2059590

Ion Ratio Lower Upper

128 100

127 13.3 10.3 15.5

129 11.3 8.7 13.1

