

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092823\
 Data File : VX037945.D
 Acq On : 28 Sep 2023 12:37
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 29 04:24:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X092723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Sep 28 06:08:32 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 09/29/2023
 Supervised By :Semsettin Yesilyurt 09/29/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.513	128	0m	30.000	ug/l	0.61
28) 1,4-Difluorobenzene	6.763	114	372260	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	356397	30.000	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	5.958	65	130794m	0.000	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	0.000%#
60) 4-Bromofluorobenzene	11.079	95	166012	28.574	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.233%
63) Toluene-d8	8.646	98	448906	29.178	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.267%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092823\
Data File : VX037945.D
Acq On : 28 Sep 2023 12:37
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/29/2023
Supervised By :Semsettin Yesilyurt 09/29/2023

Quant Time: Sep 29 04:24:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X092723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Sep 28 06:08:32 2023
Response via : Initial Calibration

