

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092823\
 Data File : VX037958.D
 Acq On : 28 Sep 2023 18:30
 Operator : JC/MD
 Sample : 04662-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-560-562

Quant Time: Sep 29 05:49:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	261782	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	446143	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	434733	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	228694	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162476	47.239	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.480%
35) Dibromofluoromethane	5.391	113	141176	47.293	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.580%
50) Toluene-d8	8.647	98	535868	48.312	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.620%
62) 4-Bromofluorobenzene	11.079	95	206511	51.237	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092823\
Data File : VX037958.D
Acq On : 28 Sep 2023 18:30
Operator : JC/MD
Sample : 04662-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-560-562

Quant Time: Sep 29 05:49:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

