

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092823\
 Data File : VX037959.D
 Acq On : 28 Sep 2023 18:54
 Operator : JC/MD
 Sample : 04667-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FRAC-TANK-A2970

Quant Time: Sep 29 05:50:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	217719	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	370803	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	189506	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	134546	47.036	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.080%
35) Dibromofluoromethane	5.391	113	115744	46.651	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.300%
50) Toluene-d8	8.646	98	452767	49.114	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.220%
62) 4-Bromofluorobenzene	11.079	95	171728	51.264	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.520%
Target Compounds						
16) Acetone	2.379	43	3566	2.591	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092823\
Data File : VX037959.D
Acq On : 28 Sep 2023 18:54
Operator : JC/MD
Sample : 04667-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-A2970

Quant Time: Sep 29 05:50:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

