

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VOX092920\
 Data File : VX018651.D
 Acq On : 29 Sep 2020 12:37
 Operator : JC/SP
 Sample : L4135-14
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BG486

Quant Time: Oct 05 02:07:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 30 04:47:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	266886	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	250127	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	118882	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	62430	4.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.40%
7) Chloroethane-d5	1.70	69	54474	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.60%
11) 1,1-Dichloroethene-d2	2.34	63	110975	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	4.55	46	190388	53.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.92%
24) Chloroform-d	5.14	84	154486	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
26) 1,2-Dichloroethane-d4	6.03	65	87745	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
32) Benzene-d6	6.05	84	289217	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
36) 1,2-Dichloropropane-d6	7.37	67	91513	5.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.00%
41) Toluene-d8	8.70	98	266315	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
45) trans-1,3-Dichloropropene-	8.99	79	36998	4.73	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.60%
48) 2-Hexanone-d5	9.43	63	153092	54.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.28%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	67675	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.40%
65) 1,2-Dichlorobenzene-d4	12.36	152	100948	5.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.00%

Target Compounds					Ovalue
53) Chlorobenzene	10.12	112	3076	0.089 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092920\
Data File : VX018651.D
Acq On : 29 Sep 2020 12:37
Operator : JC/SP
Sample : L4135-14
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG486

Quant Time: Oct 05 02:07:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 30 04:47:32 2020
Response via : Initial Calibration

