

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VO092920\
 Data File : VX018660.D
 Acq On : 29 Sep 2020 16:15
 Operator : JC/SP
 Sample : L4143-19
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BG0F9

Quant Time: Sep 30 06:35:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 30 04:47:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	244607	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	231101	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	103370	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	49572	3.87	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.40%
7) Chloroethane-d5	1.70	69	45714	4.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.80%
11) 1,1-Dichloroethene-d2	2.34	63	93689	3.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.20%
20) 2-Butanone-d5	4.56	46	172276	52.78	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.56%
24) Chloroform-d	5.14	84	136014	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
26) 1,2-Dichloroethane-d4	6.03	65	77733	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
32) Benzene-d6	6.05	84	240698	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
36) 1,2-Dichloropropane-d6	7.37	67	75687	4.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.20%
41) Toluene-d8	8.70	98	215458	4.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.20%
45) trans-1,3-Dichloropropene-	8.99	79	31360	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
48) 2-Hexanone-d5	9.43	63	132487	51.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.36%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	58160	4.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.20%
65) 1,2-Dichlorobenzene-d4	12.36	152	83731	4.96	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.20%

Target Compounds					Ovalue
14) Carbon disulfide	2.55	76	4497	0.124 ug/L	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092920\
Data File : VX018660.D
Acq On : 29 Sep 2020 16:15
Operator : JC/SP
Sample : L4143-19
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG0F9

Quant Time: Sep 30 06:35:53 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 30 04:47:32 2020
Response via : Initial Calibration

