


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092920\
Data File : VX018679.D
Acq On : 30 Sep 2020 00:43
Operator : JC/SP
Sample : L4135-06DL 40X
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG480DL

Manual Integrations
APPROVED

MMDadoda
10/1/2020 10:07:16 AM

Quant Time: Sep 30 08:36:25 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 30 07:04:14 2020
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092920\
Data File : VX018679.D
Acq On : 30 Sep 2020 00:43
Operator : JC/SP
Sample : L4135-06DL 40X
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG480DL

Manual Integrations
APPROVED

MMDadoda
10/1/2020 10:07:16 AM

Quant Time: Sep 30 08:31:31 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 30 07:04:14 2020
Response via : Initial Calibration



(8) Chloroethane (T)

1.605min (-0.110) 0.04ug/L

response 279

Ion	Exp%	Act%
64.10	100	100
66.05	31.80	39.37
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092920\
Data File : VX018679.D
Acq On : 30 Sep 2020 00:43
Operator : JC/SP
Sample : L4135-06DL 40X
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG480DL

Manual Integrations
APPROVED

MMDadoda
10/1/2020 10:07:16 AM

Quant Time: Sep 30 08:31:31 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 30 07:04:14 2020
Response via : Initial Calibration



TIC: VX018679.D

(8) Chloroethane (T)

1.721min (+0.006) 0.22ug/L m

response 1452

Ion	Exp%	Act%
64.10	100	100
66.05	31.80	45.25#
0.00	0.00	0.00
0.00	0.00	0.00

7 MB
10/06/20

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092920\
 Data File : VX018679.D
 Acq On : 30 Sep 2020 00:43
 Operator : JC/SP
 Sample : L4135-06DL 40X
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BG480DL

Manual Integrations
 APPROVED

MMDadoda
 10/1/2020 10:07:16 AM

Quant Time: Sep 30 08:36:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 30 07:04:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	223014	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	221181	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	97526	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	52601	4.51	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.20%
7) Chloroethane-d5	1.70	69	46706	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.00%
11) 1,1-Dichloroethene-d2	2.34	63	98987	3.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.80%
20) 2-Butanone-d5	4.56	46	166396	55.92	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.84%
24) Chloroform-d	5.14	84	135414	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	6.03	65	79163	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
32) Benzene-d6	6.05	84	246525	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	7.37	67	76196	4.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.00%
41) Toluene-d8	8.70	98	219723	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.60%
45) trans-1,3-Dichloropropene-	8.99	79	30501	4.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.20%
48) 2-Hexanone-d5	9.43	63	128247	51.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.52%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	60157	5.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.80%
65) 1,2-Dichlorobenzene-d4	12.36	152	84552	5.31	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.20%

Target Compounds

					Ovalue
8) Chloroethane	1.72	64	1452m	0.219	ug/L
16) Methylene chloride	2.84	84	4618	0.310	ug/L
33) Benzene	6.12	78	5224	0.123	ug/L
53) Chlorobenzene	10.12	112	307716	10.113	ug/L
58) Isopropylbenzene	11.00	105	17392	0.358	ug/L
64) 1,4-Dichlorobenzene	12.08	146	3598	0.168	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed