

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
 Data File : VX031776.D
 Acq On : 29 Sep 2022 16:19
 Operator : JC/MD
 Sample : N4860-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-07

Quant Time: Sep 30 04:34:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	55305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	95430	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	84685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66643	56.776	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.560%
35) Dibromofluoromethane	5.385	113	57556	50.618	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.240%
50) Toluene-d8	8.647	98	160759	38.997	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	78.000%#
62) 4-Bromofluorobenzene	11.079	95	48224	31.917	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	63.840%#
Target Compounds						Qvalue
73) Isopropylbenzene	10.963	105	2526	1.021	ug/l	96
83) tert-Butylbenzene	11.713	119	1533	0.758	ug/l	80
85) sec-Butylbenzene	11.890	105	3098	1.205	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
Data File : VX031776.D
Acq On : 29 Sep 2022 16:19
Operator : JC/MD
Sample : N4860-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-07

Quant Time: Sep 30 04:34:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

