

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092922\  
 Data File : VX031777.D  
 Acq On : 29 Sep 2022 16:42  
 Operator : JC/MD  
 Sample : N4860-10  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 EB-2022-09-27

Quant Time: Sep 30 04:34:35 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092622W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Sep 27 01:44:18 2022  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.550  | 168  | 72904    | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.763  | 114  | 110995   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.055 | 117  | 120789   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.024 | 152  | 85031    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| System Monitoring Compounds |        |       |          |          |      |         |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 71767    | 46.381   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 92.760% |
| 35) Dibromofluoromethane    | 5.385  | 113   | 63022    | 47.653   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 95.300% |
| 50) Toluene-d8              | 8.653  | 98    | 221889   | 46.278   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 92.560% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 82513    | 46.953   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =    | 93.900% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092922\  
Data File : VX031777.D  
Acq On : 29 Sep 2022 16:42  
Operator : JC/MD  
Sample : N4860-10  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
EB-2022-09-27

Quant Time: Sep 30 04:34:35 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092622W.M  
Quant Title : SW846 8260  
QLast Update : Tue Sep 27 01:44:18 2022  
Response via : Initial Calibration

