

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
 Data File : VX031795.D
 Acq On : 30 Sep 2022 00:49
 Operator : JC/MD
 Sample : N4864-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2022-09-27

Quant Time: Sep 30 04:46:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	112544	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120946	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71886	46.602	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.200%
35) Dibromofluoromethane	5.385	113	62557	46.650	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.300%
50) Toluene-d8	8.647	98	223853	46.045	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.080%
62) 4-Bromofluorobenzene	11.079	95	79073	44.376	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.760%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
Data File : VX031795.D
Acq On : 30 Sep 2022 00:49
Operator : JC/MD
Sample : N4864-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2022-09-27

Quant Time: Sep 30 04:46:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

