

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
 Data File : VX031796.D
 Acq On : 30 Sep 2022 01:12
 Operator : JC/MD
 Sample : N4864-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Sep 30 06:25:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	71165	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	112177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72283	47.856	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.720%
35) Dibromofluoromethane	5.385	113	62324	46.628	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.260%
50) Toluene-d8	8.653	98	218976	45.189	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.380%
62) 4-Bromofluorobenzene	11.079	95	77947	43.888	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.780%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
Data File : VX031796.D
Acq On : 30 Sep 2022 01:12
Operator : JC/MD
Sample : N4864-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Sep 30 06:25:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

