

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
 Data File : VX031797.D
 Acq On : 30 Sep 2022 04:11
 Operator : JC/MD
 Sample : N4864-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Sep 30 06:26:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69923	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.751	114	109637	50.000	ug/l	-0.01
63) Chlorobenzene-d5	10.055	117	118262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76697	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	68897	46.425	ug/l	-0.01
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.840%		
35) Dibromofluoromethane	5.379	113	59944	45.887	ug/l	-0.01
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.780%		
50) Toluene-d8	8.647	98	205364	43.362	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	86.720%		
62) 4-Bromofluorobenzene	11.079	95	74781	43.080	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.160%		
Target Compounds						Qvalue
78) n-propylbenzene	11.305	91	1661	0.301	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
Data File : VX031797.D
Acq On : 30 Sep 2022 04:11
Operator : JC/MD
Sample : N4864-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Sep 30 06:26:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

