

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
 Data File : VX031800.D
 Acq On : 30 Sep 2022 05:20
 Operator : JC/MD
 Sample : N4864-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12

Quant Time: Sep 30 06:27:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	70640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	110792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119368	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77092	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70347	46.921	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.840%
35) Dibromofluoromethane	5.385	113	60830	46.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	8.647	98	211003	44.088	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.180%
62) 4-Bromofluorobenzene	11.079	95	74970	42.739	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092922\
Data File : VX031800.D
Acq On : 30 Sep 2022 05:20
Operator : JC/MD
Sample : N4864-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12

Quant Time: Sep 30 06:27:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

