

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092923\
 Data File : VX037968.D
 Acq On : 29 Sep 2023 12:18
 Operator : JC/MD
 Sample : 04656-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-105D

Quant Time: Sep 29 23:02:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	212510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	358151	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	342665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176918	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	130349	46.686	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.380%
35) Dibromofluoromethane	5.391	113	110720	46.203	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.647	98	432189	48.538	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.080%
62) 4-Bromofluorobenzene	11.079	95	160274	49.535	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092923\
Data File : VX037968.D
Acq On : 29 Sep 2023 12:18
Operator : JC/MD
Sample : 04656-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-105D

Quant Time: Sep 29 23:02:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

