

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092923\
 Data File : VX037971.D
 Acq On : 29 Sep 2023 13:27
 Operator : JC/MD
 Sample : 04656-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-106I

Quant Time: Sep 29 23:04:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	228988	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	390551	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	384829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	206139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140487	46.696	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.400%
35) Dibromofluoromethane	5.385	113	121508	46.498	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.000%
50) Toluene-d8	8.647	98	484539	49.903	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.800%
62) 4-Bromofluorobenzene	11.079	95	183023	51.873	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.740%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.495	96	5116	1.682	ug/l	81
44) Trichloroethene	7.141	130	1271	0.412	ug/l	72
64) Tetrachloroethene	9.275	164	1049	0.400	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092923\
Data File : VX037971.D
Acq On : 29 Sep 2023 13:27
Operator : JC/MD
Sample : 04656-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-106I

Quant Time: Sep 29 23:04:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

