

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092923\
 Data File : VX037979.D
 Acq On : 29 Sep 2023 16:33
 Operator : JC/MD
 Sample : 04656-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-104D

Quant Time: Sep 29 23:07:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	217918	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	367013	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	351141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	182260	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135694	47.394	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.780%
35) Dibromofluoromethane	5.391	113	114894	46.787	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.580%
50) Toluene-d8	8.647	98	448965	49.205	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.400%
62) 4-Bromofluorobenzene	11.079	95	163055	49.177	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092923\
Data File : VX037979.D
Acq On : 29 Sep 2023 16:33
Operator : JC/MD
Sample : 04656-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-104D

Quant Time: Sep 29 23:07:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

